

**U.S. DEPARTMENT OF AGRICULTURE  
Grain Inspection, Packers and Stockyards Administration  
Federal Grain Inspection Services**

CERTIFICATE NO. GIPSA 2007-009

CERTIFICATE OF PERFORMANCE

QuickStix™ Kit for mCry3A (Agrisure™ RW) Corn Bulk Grain, Qualitative Lateral Flow Strip for Corn

For: EnviroLogix, Inc.  
Method: Lateral Flow Strip

Submitted by:  
EnviroLogix, Inc.  
500 Riverside Industrial Pkwy.  
Portland, ME 04103-1418  
Tel: (207) 797-0300  
Contact: Ms. Joan Lawton

Standard Features and Options

<u>Model</u>	<u>Event(s) Detected</u>	<u>Sample Preparation</u>	<u>Extraction</u>
Part #AS 037 BG QuickStix™ Kit for mCry3A (Agrisure™ RW) Corn Bulk Grain	Genetic Events expressing the mCry3A Protein	Grind	Tap Water

Test kits must be operated according to the manufacturers instructions.

This qualitative lateral flow strip test kit underwent an initial verification of performance under the authority of Section 7B (c) of the United States Grain Standards Act, as amended, and was found to meet all test performance criteria as defined in "Performance Verification of Qualitative Mycotoxin and Biotech Rapid Test Kits" March 2004 version. Evaluation tests that were passed are summarized in Attachment 1.

For further information, contact:

USDA, Grain Inspection, Packers and Stockyards Administration  
Technical Services Division  
Quality Systems & Services Unit  
10383 N. Ambassador Drive  
Kansas City, Missouri 64153-1394 Tel: (816) 891-0401

  
Director  
Technical Services Division

Date: 8/7/2007

*Certificate Expires Three Years from the Date Signed*

Note: The mention of firm name or trade products does not imply that they are endorsed or recommended by the United States Department of Agriculture over other firms or similar products not mentioned.

Type Evaluation  
Certificate No. GIPSA 2007-009

ATTACHMENT 1

Manufacturer: EnviroLogix, Inc.  
500 Riverside Industrial Pkwy.  
Portland, ME 04103-1418  
(207) 797-0300

TEST 1: CAPABILITY OF ANALYZING FOR THE mCry3A PROTEIN.

Data submitted by the manufacturer showed that the test kit was capable of detecting the mCry3A Protein at a threshold of 6 mCry3A kernels in 800 kernels of corn.

TEST 2: TEMPERATURE SENSIVITY

This requirement is waived for lateral flow strip test kits.

TEST 3: REPRODUCIBILITY

Data submitted by test kit manufacturer supported reproducibility claims.

TEST 4: RUGGEDNESS

Data submitted by the manufacturer supported ruggedness claims.

TEST 5: GIPSA EVALUATION

Data generated by GIPSA staff showed that the test kit was capable of detecting the mCry3A Protein at a threshold of 6 mCry3A kernels in 800 kernels of corn.