

# QuickStix™ Combo Kit for Multi-Trait Testing

## Cry1Ab/2Ae/2m/*bar* Cotton Leaf & Seed

Catalog Number AS 059 LS

Part #11573, 11574(S)

### Highlights:

- Recognizes Cry1Ab, Cry2Ae, 2mEPSPS, and LibertyLink® traits
- Positive results in 10 minutes or less

### Contents of Kit:

- 100 QuickStix Combo Strips packed in two moisture-resistant canisters
- 100 Disposable Tissue Extractors (each consisting of a tube and pestle, with punch cap)
- EB2 Extraction Buffer

### Items Not Provided:

- Repeating pipetter or other means of dispensing 0.5 mL

Contact EnviroLogix to order bulk-packaged kits. Bulk kits include EB2 Extraction Buffer Concentrate. To prepare 1 liter, mix 50 mL 20X Concentrate with 950 mL of distilled or deionized water. Store refrigerated when not in use; allow to come to room temperature before using.



Crush single seed



## Intended Use

The EnviroLogix QuickStix Combo Kit for Multi-Trait Testing is designed to detect the presence of Cry1Ab & Cry2Ae specific to TwinLink® (Event T304-40 X GHB119), along with 2mEPSPS and PAT/*bar* proteins, in separate regions of the same strip, as expressed in genetically modified cotton leaf or single seed tissue.

## How the Test Works

Cotton plants and seeds that have been genetically modified with stacked genes express transgenic proteins in their tissue. To detect these proteins with the QuickStix Combo Strips, the sample must be extracted and the proteins solubilized in the Extraction Buffer provided.

Each Combo Strip has an absorbent pad at each end. The protective tape with the arrow indicates which end of the strip to insert into the sample. The sample travels up the membrane strip and is absorbed into the larger pad at the top of the strip. The portion of the strip between the protective tape and the absorbent pad at the top of the strip is used to view the reactions as described under “Interpreting the Results”.

## Sample Preparation

*Note: If Extraction Buffer has been refrigerated, allow it to warm up to room temperature before preparing samples.*

### To extract cotton seed:

1. Crush a single cotton seed (*Suggestion: use pliers with seed in microcentrifuge tube or resealable plastic bag*). Transfer to an extraction tube marked with sample identification. Note: Complete crushing of seed improves extraction efficiency and test performance.
2. Add **0.5 mL** Extraction Buffer.
3. Close the tube cap securely. Shake the tube vigorously for 20 to 30 seconds, using an **up-and-down motion**, ensuring that the crushed seed and buffer are **well mixed**. Allow the solid material to settle to the bottom of the tube. The extract takes on a yellow to brown opaque color when the samples are prepared properly.
4. Use caution to prevent sample-to-sample cross-contamination with plant tissue, fluids, crushing equipment (*pliers*) or disposables. Be sure to use a new tube for each sample tested.

### To extract cotton leaf:

1. Sandwich a section of leaf tissue between the cap and body of the Disposable Tissue Extractor tube; snap two circular tissue punches by closing the cap. Push the leaf punches down into the tapered bottom of the tube with the pestle. Sample identification should be marked on the tube with a waterproof marker.
2. Insert the pestle into the tube and grind the tissue by rotating the pestle against the sides of the tube with twisting motions. Continue this process for 20 to 30 seconds, or until the leaf tissue is well ground.



*Obtain Leaf Tissue*



*Grind Tissue*



*Insert QuickStix*

*Any clearly discernible pink  
Test Line is considered positive*

3. Uncap the bottle of Extraction Buffer; carefully add **0.5 mL** of Buffer into the tube containing cotton leaf.
4. Repeat the grinding step to mix tissue with Extraction Buffer. Dispose of the pestle (do not re-use pestles on more than one sample).
5. Repeat the protocol for each sample to be tested, using a new tube and pestle for each. Use caution to prevent sample-to-sample cross-contamination with plant tissue, fluids, or disposables.

## How to Run the QuickStix Strip Test

1. Allow refrigerated canisters to come to room temperature before opening. Remove the QuickStix Strips to be used. Avoid bending the strips. Reseal the canister immediately.
2. Place the strip into the extract. The sample will travel up the strip. Use a rack to support multiple tubes if needed.
3. Allow the strip to develop for 10 minutes before making final assay interpretations. Positive sample results may become obvious much more quickly.
4. To retain the strip, cut off and discard the bottom section of the strip covered by the arrow tape.

## Interpreting the Results

Development of the Control Line within 10 minutes indicates that the strip has functioned properly. Any strip that does not develop a Control Line should be discarded and the sample re-tested using another strip.

**One Line** – If the extract is from a negative sample, containing none of the target proteins, the strip will only show the Control Line, about 4 mm below the top pad.

**Two to Five Lines** – If the extract is from a sample containing Cry1Ab, Cry2Ae, 2mEPS (GlyTol<sup>®</sup>), and/or PAT/*bar* (LibertyLink) proteins, the appropriate test line(s) will appear. Please refer to the interpretation guide (next page).

Please note that for ease of interpretation, the labels on each strip list the proteins in the order that their Test Lines appear, bottom (arrow end) to top (lot number end).

## Kit Storage

This Kit can be stored at room temperature, or refrigerated for a longer shelf life. Please note the shelf life on the kit label for each storage temperature. The kit may be used in field applications; however, prolonged exposure to high temperatures may adversely affect the test results. Do not open the desiccated canister until ready to use the test strips.

## Precautions and Notes

- This kit is designed for screening for presence or absence only and is not meant to be quantitative.
- This product is currently not applicable for use in any other crop than cotton.
- As with all tests, it is recommended that results be confirmed by an alternate method when necessary.

- The assay has been optimized to be used with the protocol and with the buffer provided in the kit. Deviation from this protocol may invalidate the results of the test.
- A negative result with this kit does not mean that the sampled tissue has not been otherwise genetically modified.
- A strong positive result may safely be interpreted in as little as 2 minutes after sample addition. It is not safe, however, to conclude that a sample is negative before a full 10 minutes has elapsed.
- Protect all components from hot or cold extremes of temperature when not in use. Do not leave in direct sunlight or in vehicle.



Interpretation Guide				
	+ <i>bar</i> , +2m, +Cry2Ae, +Cry1Ab (e.g. GLT: GlyTol/LibertyLink/TwinLink)	- <i>bar</i> , -2m, -Cry2Ae, -Cry1Ab (e.g. varieties that do not contain these four events or non-GMO)	+ <i>bar</i> , -2m, -Cry2Ae, -Cry1Ab (e.g. LibertyLink, BollgardII/LL, BollgardIII/RR)	+ <i>bar</i> , +2m, -Cry2Ae, -Cry1Ab (e.g. GlyTol/LibertyLink)
Control Line	✓	✓	✓	✓
Cry1Ab Test Line	✓			
Cry2Ae Test Line	✓			
2mEPSPS Test Line	✓			✓
PAT/ <i>bar</i> Test Line	✓		✓	✓

✓ Indicates where a test line will appear (positive result for that specific trait)  
 Indicates no test line (negative result for that specific trait)

Please note that for ease of interpretation, the labels on each strip list the proteins in the order that their Test Lines appear, bottom (arrow end) to top (lot number end).



**For Technical Support  
Contact Us At:**

**EnviroLogix**  
500 Riverside Industrial  
Parkway  
Portland, ME 04103-1486  
USA

**Tel: (207) 797-0300**  
**Toll Free: 866-408-4597**  
**Fax: (207) 797-7533**

e-mail:  
***info@envirologix.com***

website:  
***www.envirologix.com***



**LIMITED WARRANTY**

EnviroLogix Inc. (“EnviroLogix”) warrants the products sold hereunder (“the Products”) against defects in materials and workmanship when used in accordance with the applicable instructions for a period of one year from the date of shipment of the products or if shorter, for a period not to extend beyond a product’s printed expiration date. If the Products do not conform to this Limited Warranty and the customer notifies EnviroLogix in writing of such defects during the warranty period, including an offer by the customer to return the Products to EnviroLogix for evaluation, EnviroLogix will repair or replace, at its option, any product or part thereof that proves defective in materials or workmanship within the warranty period.

**NEITHER ENVIROLOGIX NOR MONSANTO MAKES ANY OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.** The warranty provided herein and the data, specifications and descriptions of EnviroLogix products appearing in EnviroLogix published catalogues and product literature are EnviroLogix’ sole representations concerning the Products and warranty. No other statements or representations, written or oral, by EnviroLogix’ employees, agents or representatives, except written statements signed by a duly authorized officer of EnviroLogix Inc., are authorized; they should not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

EnviroLogix does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by EnviroLogix, or against damages resulting from such non-EnviroLogix made products or components. EnviroLogix passes on to customer the warranty it received (if any) from the maker thereof of such non-EnviroLogix made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by EnviroLogix.

**THIS WARRANTY IS EXCLUSIVE.** The sole and exclusive obligation of EnviroLogix shall be to repair or replace the defective Products in the manner and for the period provided above. EnviroLogix shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall EnviroLogix be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of EnviroLogix with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

EnviroLogix has developed this kit using proprietary reagents as well as reagents licensed from Monsanto Company.

*LibertyLink, TwinLink and GlyTol are registered trademarks of Bayer  
Bollgard II is a registered trademark of Monsanto Technology, LLC  
EnviroLogix, the EnviroLogix logo, and QuickStix are trademarks of EnviroLogix Inc.*

© EnviroLogix 2015