

*Designed for use with
DNable Molecular Detection Kits
for Salmonella and Salmonella
Plus, and an appropriate Sample
Extraction Set*

Contents

- Dried positive control, 2 tubes

Catalog No. ACC 226

Part #11768

Intended Use

This positive control is intended for the user verification of the DNable Molecular Detection Kits for *Salmonella* (Catalog No. DF 026, Part #11716) and *Salmonella* Plus (Cat. No. DF-126, Part #12010). This is a non-biologically active, non-hazardous positive control provided in dried form.

Intended User

Customers using the DNable Molecular Detection Kits for *Salmonella* and *Salmonella* Plus who wish to incorporate positive and negative (optional) control(s). Training in and support of the use of the DNable *Salmonella* assays is available through EnviroLogix. It is designed for use by personnel with appropriate training in handling human pathogens, and in Microbiology and Molecular Assay techniques. Users should define the frequency of control testing based on their facility's Quality Policies.

Components

A. Positive control, two tubes

B. DNase-free water, 1 tube

Procedures

Positive Control

1. Suspend one tube of positive control in 65 µL of DNase-free water. Vortex briefly for uniform resuspension. Suspended control may be stored at 2-8°C for up to 2 weeks. Do not put the suspended positive control vial back in the desiccant pouch.
2. Transfer 5 µL to a green RB1 tube from either of the DNable Molecular Detection Kits for *Salmonella*.
3. Follow instructions in the Molecular Detection Kit for amplification and detection.

Negative Control from Sample Extraction Set 1 or 2

1. Transfer 5 µL of MB2 or MB3 from DNable Sample Extraction Set 1 or 2, respectively to a green RB1 tube from the DNable Molecular Detection Kit for *Salmonella*.
2. Follow instructions in the Molecular Detection Kit for amplification and detection.

Interpretation and Troubleshooting

- Positive amplification results should be obtained for the positive control. Negative results may indicate an issue with reagents or procedures. Users are advised to repeat procedures or contact EnviroLogix Technical support.
- Negative controls should be interpreted as negative. Positive results may indicate contamination. Replicate analyses may be attempted or EnviroLogix Technical Support may be consulted.



For Technical Support
Contact Us At:

EnviroLogix
500 Riverside Industrial Parkway
Portland, ME 04103-1486 USA
Tel: (207) 797-0300
Toll Free: 866-408-4597
Fax: (207) 797-7533

e-mail: dnable@envirologix.com

Website: www.envirologix.com



DNAble

Salmonella Positive Control Accessory Set

Part #11768

LIMITED WARRANTY

EnviroLogix Inc. ("EnviroLogix") warrants the products sold hereunder ("the Products") against defects in materials and workmanship when used in accordance with the applicable instructions for a period not to extend beyond a product's printed expiration date. If the Products do not conform to this Limited Warranty and the customer notifies EnviroLogix in writing of such defects during the warranty period, including an offer by the customer to return the Products to EnviroLogix for evaluation, EnviroLogix will repair or replace, at its option, any product or part thereof that proves defective in materials or workmanship within the warranty period.

ENVIROLOGIX MAKES NO OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. The warranty provided herein and the data, specifications and descriptions of EnviroLogix products appearing in EnviroLogix published catalogues and product literatures are EnviroLogix' sole representations concerning the Products and warranty. No other statements or representations, written or oral, by EnviroLogix' employees, agents or representatives, except written statements signed by a duly authorized officer of EnviroLogix Inc., are authorized; they should not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

EnviroLogix does not warrant against damages or defects arising in shipping or handling, or out of accident or improper or abnormal use of the Products; against defects in products or components not manufactured by EnviroLogix, or against damages resulting from such non-EnviroLogix made products or components. EnviroLogix passes on to customer the warranty it received (if any) from the maker thereof of such non-EnviroLogix made products or components. This warranty also does not apply to Products to which changes or modifications have been made or attempted by persons other than pursuant to written authorization by EnviroLogix.

THIS WARRANTY IS EXCLUSIVE. The sole and exclusive obligation of EnviroLogix shall be to repair or replace the defective Products in the manner and for the period provided above. EnviroLogix shall not have any other obligation with respect to the Products or any part thereof, whether based on contract, tort, strict liability or otherwise. Under no circumstances, whether based on this Limited Warranty or otherwise, shall EnviroLogix be liable for incidental, special, or consequential damages.

This Limited Warranty states the entire obligation of EnviroLogix with respect to the Products. If any part of this Limited Warranty is determined to be void or illegal, the remainder shall remain in full force and effect.

EnviroLogix, the EnviroLogix logo, DNAble, and the DNAble logo are trademarks of EnviroLogix Inc.

©EnviroLogix 2021